

VetScreen Influenza A/H1N1

User Manual
Cat. No. V03-01-1131

One-step reverse transcription PCR for gel-based
detection of influenza type A/H1N1 subtype
in veterinary and environmental samples

Includes main components for 50 reactions



Rev. 0
June 2010

HAI KANG LIFE CORPORATION LIMITED

PLEASE READ THROUGH THE ENTIRE PROTOCOL BEFORE
STARTING.

1. KIT COMPONENTS

DNA Amplification Reagents

- 2 x 500 µl H1N1 Mastermix (store at -20°C)
- 2 x 25 µl Positive Control (store at -20°C)
- 1 x 25 µl Taq Polymerase (store at -20°C)

Storage Conditions

Store Mastermix at -20°C. Thaw frozen reagents just before use. Mix reagents thoroughly (do not vortex mastermix containing enzyme).

Note:

Please aliquot the Mastermix into appropriate volume according to your test frequency, in order to minimize repeated freeze and thaw cycles. Frequent thawing and freezing may inactivate some kit components.

2. PROCEDURES

Set-up

- In addition to the RNA obtained from the test samples, each experiment requires a positive and negative (water) control.
- Set up PCR components according to the table below:

Components	Volume per reaction
Mastermix	19.5 µl
Taq Polymerase	0.5 µl
Sample RNA	5 µl
Total Volume	25 µl

Note:

- 5 µl of the provided Positive Control can be used to monitor the success of amplification.
- Keep RNA samples on ice throughout experiment.
- It is advisable to run samples in duplicate to ensure reliability of results.
- Spiking Control: 1 µl Positive Control can be spiked into test sample to check whether the test sample contains PCR inhibitory substances.

Cycling conditions:

1 Cycle	42°C	30 Minutes
	95°C	10 Minutes
40 Cycles	95°C	30 Seconds
	55°C	30 Seconds
	72°C	30 Seconds

3. DATA ANALYSIS AND INTERPRETATION

Expected PCR product size: 190 base pairs

Spiking Control

Negative PCR result may be due to a few scenarios: 1. absence of detected sequence in the sample; 2. presence of detected sequence below limit of detection; 3. presence of PCR inhibitory substances. The purpose of spiking control is to verify whether the test sample contains substances, which may affect PCR reactions. If no band is visible on agarose gel when the test sample is spiked with the positive control, the test sample is highly likely to contain PCR inhibitory substances, and the result should NOT be taken as negative. Repeated extraction and PCR of the sample will be required.

If you require more detailed analysis information, please contact Hai Kang Life Corporation Limited for technical assistance.

4. TECHNICAL ASSISTANCE

Our technical staff will provide technical assistance you may need in using this kit. Simply call +(852) 2111 2123 during office hours:

Monday – Friday: 9:00am to 5:30pm

Saturday: 9:00am to 1:00pm

A recorded message (in English, Cantonese or Putonghua) may be left outside office hours. Alternatively, you may contact our technical staff by fax or email.

Fax: +(852) 2111 9762

Email: technical@haikanglife.com

5. WARRANTIES AND LIABILITIES

Hai Kang Life Corporation Limited warrants the products manufactured by it are free of defects in materials and workmanship when used in accordance with the applicable instructions for a period equal or shorter of one year from the date of shipment of the product(s) or the expiration date marked on the product packaging under the storage conditions recommended in the instructions and/or on the package. Application protocols published by Hai Kang Life Corporation Limited are intended to be only guidelines for the buyers of the products. Buyers are expected to validate the kit to their individual application. Hai Kang Life Corporation Limited makes no other warranty, expressed or implied. There is no warranty of merchantability or fitness for a particular purpose.

The sole obligation of Hai Kang Life Corporation Limited with regard to the foregoing warranties shall be, at its option, to either replace or refund the purchase price of the product(s) or part thereof that proves defective in materials or workmanship within the warranty period, provided the customer notifies Hai Kang Life Corporation Limited promptly of any such defect. Hai Kang Life Corporation Limited shall not be liable for any defect, indirect or consequential damages resulting from economic loss or property damages sustained by the buyer or any customer from the user of the product(s).

Notice

The PCR process is covered by US patents owned by F. Hoffman-La Roche, Inc. Use of the PCR process requires a license. Nothing in this publication should be construed as an authorisation or implicit license to practice PCR under any patents held by F. Hoffman-La Roche, Inc.