

CHROMOSOMAL ANEUPLOIDIES BY QUANTITATIVE FLUORESCENT PCR ANALYSIS (for Amniotic Fluid)

Extra Primer Set

Cat. No. C02-01-1122



Rev. 0
July 2007

HAI KANG LIFE CORPORATION LIMITED

Notice

The PCR process is covered by US patents owned by F. Hoffman-La Roche, Inc. Use of the PCR process requires a license. Nothing in this publication should be construed as an authorisation or implicit license to practice PCR under any patents held by F. Hoffman-La Roche, Inc.

1. KIT COMPONENTS

The kit contains the following components sufficient to process approximately 25 amniotic fluid samples and should be stored at -20°C. Repeated freezing and thawing cycles will inactivate some components, therefore, keeping the components in small aliquots is highly recommended.

- PCR mastermix (1 x 530 µl)
- Primer 21-4 (1 x 50 µl)
- Primer 21-5 (1 x 50 µl)
- Primer 18-3 (1 x 50 µl)
- Primer 13-3 (1 x 50 µl)

Note, the following additional materials are also required.

- DNA extraction kit for amniotic fluid (QIAamp DNA Blood Kit (cat No. 51104) or other method giving DNA of similar PCR quality)
- Gene Scan TAMRA 500 size standard (ABI cat. No. 401733)
- ABI Hi-Di™ formamide

2. PROCEDURE

2.1 DNA Extraction From Amniotic Fluid

Method

1. Centrifuge 1.5 ml of amniotic fluid for 10 min at 4°C at 1000 *g*, then discard the supernatant.
2. Add 200 µl of ice-cold 1 X PBS to resuspend the pellet.
3. Follow instructions of Qiagen Blood Protocol with exception of final elution step: Add 100µl **AE** or distilled water to the QIAamp spin column, incubate for at least 5 min at room temperature. Centrifuge at 6000 *g* at 25°C for 1 min to collect DNA from the sample.

2.2 Quantitative fluorescent PCR (QF-PCR)

Method

1. For EACH PCR reaction, mix the following components in a total volume of 25 µl. One negative control (reaction without template) should be run with each set of experiments.

Components	Vol. for 1X
dd water	17.7 µl – n
PCR mastermix	5.3 µl
Primer	2 µl
Sample DNA	n
Total volume	25 µl

Note: For amniotic fluid extracted by the QIAamp DNA Blood Kit, 5 µl sample DNA is sufficient for PCR amplification. For other DNA extraction method, please use 50–100 ng sample DNA.

2. The PCR profile required to run the pre-natal screening reactions is optimized for PCR thermal cycler with temperature ramping rate of 3°C/s. Program these steps:

1 Cycle: 95°C, 11 min
28 Cycles: 94°C, 1 min
 59°C, 1 min
 72°C, 1 min
1 Cycle: 60°C, 45 min

3. Analysis of fluorescent PCR products.

- a. For each analysis, 4 µl of PCR product is mixed with 16 µl Hi-Di™ formamide and 1 µl of Gene Scan TAMRA 500 size standard.
- b. Heat the mixture at 95°C for 5 min.
- c. Cool the mixture immediately on ice for at least 30 s.
- d. Briefly centrifuge the mixture.
- e. Proceed to analysis using ABI PRISM Genetic Analyzer.

For detailed operation procedures on using the ABI PRISM Genetic Analyzer, please refer to the relevant User Manual provided by your supplier.

3. Data collection, analysis and interpretation

Table 1: PCR product information

Primer mix

Marker	Size of PCR Products (bp)	Color of Plot in ABI 310
21-4	163-187	Black
21-5	~200	Blue
18-3	~358	Blue
13-3	152-172	Blue

PCR products are resolved on an automated laser fluorescent sequencer e.g. ABI PRISM 310 Genetic Analyzer.

For each STR marker, a normal sample produces either a single peak (homozygous state) or two (diallelic) peaks with peak area ratio of about 1:1 (heterozygous state). Trisomy genotypes show characteristic three (triallelic) peaks of similar intensity or two (diallelic) peaks with a dosage ratio of 2:1 (measured as peak areas). Additional markers are used in case of ambiguous result or uninformative homozygous peaks. The combined heterozygosity produced by two or more STRs lowers the likelihood of uninformative results (i.e. homozygosity at the loci).

If you require more detailed analysis information please contact Hong Kong DNA Chips for technical assistance.

4. TECHNICAL ASSISTANCE

Our technical staff will provide technical assistance you may need in using this kit. Simply call +(852) 2111 2123 during office hours:

Monday – Friday: 9:00am to 5:30pm
Saturday: 9:00am to 1:00pm

A recorded message (in English, Cantonese or Putonghua) may be left outside office hours.

Alternatively, you may contact our technical staff by fax or email.

Fax: + (852) 2111 9762
Email: technical@haikanglife.com

5. WARRANTIES AND LIABILITIES

Hai Kang Life Corporation Limited warrants the products manufactured by it are free of defects in materials and workmanship when used in accordance with the applicable instructions for a period equal or shorter of one year from the date of shipment of the product(s) or the expiration date marked on the product packaging under the storage conditions recommended in the instructions and/or on the package. Application protocols published by Hai Kang Life Corporation Limited are intended to be only guidelines for the buyers of the products. Buyers are expected to validate the kit to their individual application. Hai Kang Life Corporation Limited makes no other warranty, expressed or implied. There is no warranty of merchantability or fitness for a particular purpose.

The sole obligation of Hai Kang Life Corporation Limited with regard to the foregoing warranties shall be, at its option, to either replace or refund the purchase price of the product(s) or part thereof that proves defective in materials or workmanship within the warranty period, provided the customer notifies Hai Kang Life Corporation Limited promptly of any such defect. Hai Kang Life Corporation Limited shall not be liable for any defect, indirect or consequential damages resulting from economic loss or property damages sustained by the buyer or any customer from the user of the product(s).

Notice: The PCR process is covered by US patents owned by F. Hoffman-La Roche, Inc. Use of the PCR process requires a license. Nothing in this publication should be construed as an authorization or implicit license to practice PCR under any patents held by F. Hoffman-La Roche, Inc.